

**Indian Red Cross Society  
Blood Bank  
National Headquarters  
1-Red Cross Road,  
New Delhi - 110001.**

**Notice Inviting Quotation for consumables items/ Printing Material/ Badges/ Juice Drink etc.**

1. On behalf of the Secretary General, Indian Red Cross Society (Blood Centre), sealed quotation are invited under Two Bid System i.e. Technical Bid and Financial Bid from local vendors for consumable Items, kits and printing material for its Blood Bank at the above address:-

**2. Schedule of Events:**

- (a) Start Date of receiving the quotation 15-02- 2025
- (b) End Date of last submission 10-03-2025.

3.1 The sealed quotation should reach the Director, Blood Centre, Indian Red Cross Society, New Delhi-z10001 latest by 3.00 pm, on dated 10.03.2025

**3.2 Please Send**

- (a) One sealed envelope:
  - (i) The Technical bid
  - (ii) Sealed samples as per list at annexure "A" and
  - (iii) Documentary evidence of total amount of turnover being not less than Rs.49,00000/- (Rupees forty-nine lakhs only) at least 1 of the less 3 years ending on 31<sup>st</sup> March, 2024.
  - (iv) Vender register with DS&D Udyan or supplying in Govt. Hospital/ premier Health Institute.
- (b) Acceptable evidence is by way of assessment of GST.
- (c) In another sealed envelope the Financial Bid of prices valid for two years and the same quantity material is valid for second year also.
- (d) Both the sealed envelope at (a) and (b) should be put in an outer sealed envelope which should be addressed to the Director, Blood Centre at the address given in paragraph 3.1 above.

Please note that each of the envelopes should have the name, address and telephone number of the bidder.

3.3 Technical bid will be opened by the technical committee only for the bidder who have submitted the bid with the sample.

3.4 The detail technical specification may be referred for submitting the bid documents (Annexure 'A')

3.5 The inner and outer envelopes should be marked Bid FOR Consumable Items/ Printing materials/ Badges/Juice Drink in blood centre.

4.1 Technical Committee approved by the competent authority shall evaluate & decide bids are found acceptable after evaluation the equipments on demonstration and the samples of consumables/kits/reagents etc. (As per list of Annexure 'A')

4.2 The bidders whose technical bids are found acceptable by the Technical Committee shall be informed, by email.

4.3 After the opening, the financial bids shall be evaluated.

- 4.4 To ensure timely supply the 3 lowest bidder L-1 & L-2 & L3 will be short listed. L-2 will be asked to match the rates of L-1; If L2 refuses, L-3 will be asked to match the rates of L-1.
- 4.5 The order will be split among two suppliers in the ratio of 70 to L-1 and 30 L-2/3.
- 4.6 If two or more bidders become L-1 after the opening of the Financial bids and the order placed equally among the selected bidders.
5. Bids not received in sealed cover and after the due date will not be considered.
6. The Secretary General has the right to reject or accept any or all bids without assigning any reason.
7. The Supplier (the bidder(s) on whom orders are to be placed will be informed to deposit security money @ 10% of the installment amount of proposed order the security money shall be in the form of a bank draft drawn on any schedule bank in favour of the 'Indian Red Cross Society Blood Bank' New Delhi. The draft should be submitted at the time of collection of order from the society. No interest will be allowed in the above deposit. The security money will be returned to the firm after completing the order.
8. No advance payment to be give to the bidders.
9. The order of the supplier will be placed as per approved specification as per annexure 'A', Number of installment as requirement in installments.
10. If vender not supply the items as per order, security money may be forfeited and 10% penalty charges of total bills will be payable by the vender.
11. In case supply is not made in accordance with the samples provided to Red Cross Blood Centre, the order placed will stand cancelled and the security money will be forfeited.

IRCS NHQ Blood Centre

---

### Annexure – A

Indian Red Cross Society  
National Headquarters  
Blood Centre

List of following items and schedule of supply for the year 2025 – 2026 and 2026 – 2027 – reg.

Sl. No.	Name of Items	Quantity required	Supply in Installment	Specifications	Remarks
1	VDRL / RPR Kits	40000 Test	08	Antigen Kit –RPR Modified Form which contain the carbon particles to improve the Visual reading of the result.	<b>Annexure “A”</b>
2	Anti Sera	A=300,B=300,D=200, Blend=50 AB=50,H=20,A1lactin=10, Combs=10 Bovine Serum Albumin=10 nos.	03	Conforming to Quality Control Standards for Anti Sera. Anti Sera should have the titer above 1:256, Monoclonal -10ml/5ml	
3	Printer Roll for Elisa Reader	200 nos.	01	Thermal paper Roll (Good quality)	
4	Voluntary Blood Donor Forms (English)	80,000 nos. & Hindi = 5000 nos.	01	As per requirement	
5	Sticker different type (Big / Small)	30,000 nos. different color (blood gr.) 15,000 nos. different color (Compatibility)	01	As per requirement different colour and different size	
6	Distilled Water	1000 Canes	08	5 Ltr., 10 Ltr Can Ph = 7 ,TDS=less than 50PPM ,Clear, colorless Odorless and tasteless	
7	Sodium Hypochlorite	400 Canes	08	5 ltr Can 4%.The solution shall be a clear liquid free from suspended of particulate matter and shall be miscible in all proportions with distilled water, PH value greater than 11 at 20°C	
8	Fruit juice drink	50,000 nos. (180ml pack)	In installment as per requirement	Minimum 20-25% pulp it should having FSSAI Licensing Not in glass bottle	
9	Register Printing Different type	10 Nos. different type	02	As per requirement	
10	Donor Cards (Voluntary)	(50,000 nos.)	01	As per requirement	
11	Badges (Small)	20,000 nos.	01	One type badge and good quality	
12	Dustbin Plastic with lid	30 nos. 10 Ltrs.	01	First number plastic, good quality	

		15 Ltrs. 30 Ltrs.			
13	Polyester Sticker	50 Rolls	5	Good sticking power	
14	B.P. Instrument	30 Nos.	1	Dial type portable	

### ANNEXURE - A

#### **Technical Specifications of RPR (Rapid Plasma Reagin) Testing Kits:**

1. The kit should have approval of the statutory authority in its country of origin
2. In case of imported kits it should be registered and licensed under the provisions of Drugs & Cosmetics Act and rules and / or Medical Devices Rules 2017 in India.
3. In case of indigenous manufacturers should be licensed under the provisions of Drugs & Cosmetics Act and rules and or Medical Devices Rules 2017 issued by the competent authority defined under Drugs and Cosmetics Act, 1940.
4. The assay should allow for qualitative and semi quantitative determination of Reagin antibodies in serum or plasma for sero-diagnosis of syphilis based on flocculation principle using non treponemal antigens.
5. The assay should be suitable to perform with either serum or plasma
6. The assay should have sensitivity of more than or equal to 85% in primary syphilis and a specificity of more than or equal to 93%.
7. The assay should be calibrated to WHO reference serum and the same should be supported by statements in kit insert and certificate from manufacturer.
8. The test should be able to yield results within 20 minutes.
9. The pack size of RPR test kit should be less than or equal to 50 tests per kit.
10. The assay components should include positive and negative serum controls sufficient for conducting 20% of the tests (10% negative and 10% positive controls)
11. The kit should have all essential accessories required for the test such as cards, droppers, applicator, etc. in adequate quantities for the number of tests to be performed.
12. The kit should have minimum shelf-life of 5 /6<sup>th</sup> or 12 months (whichever is more) at the port of discharge of consignees.
13. The cumulative time temperature indicator technology used should be pre-qualified by WHO
14. Literature, detailing the components, methodologies, validity criteria, performance characteristics, storage conditions, manufacturing and expiry dates should be provided with each kit.